

NAPT ILC/PT Participant Tracking - Schedule & Results

Name of Participant: *ABC Calibrations*

Participant Evaluation Period: *01-Jan-07 to 31-Dec-08*

Primary Contact: *John Doe*

Appreciation Points Program: *As of 28-Nov-07 you have accumulated 46 Appreciation Points*

ILC/PT's Scheduled

ILC Number	Description	Requested Date	Tenative Start Date	Current Status
TORQ-201	Torque Wrench Set 1 - Set of Three Wrenches	28-Nov-07	01-Feb-08	Waiting for artifact
CAPACITANCEs1409	Capacitance Kit #1 GenRad 1409's	28-Nov-07	04-Apr-08	Waiting for artifact
FORCE-401	Mark-10 100lb Digital Force Gage	28-Nov-07	18-Jul-08	Waiting for artifact
THTP-104	Thread Plug Set 2 / THTP104	28-Nov-07	05-Sep-08	Waiting for artifact
SIND-100	Standard Inductor Set / NAPT-SIND-100	28-Nov-07	07-Nov-08	Waiting for artifact

ILC/PT's Participated In During Evaluation Period

ILC Number	ILC/PT Kit Description:	Block Results	Performed Date	Technician
		# of Participants		
		# of Data Points		
		Sat/UnSat Rating		
ELECTRICALs732B	10 Volt Reference / NAPT-ELECTRICALs 732B	22	05-Feb-07	John Doe
		1		
		100% Satisfactory		
GAGE-105	Gage Block Set 6 / GAGE-105	18	02-Apr-07	Jane Doe
		10		
		95.45% Satisfactory		
DMM-101	HP 34401 6 1/2 Digit Multimeter	21	04-Jun-07	Roger Doe
		10		
		100% Satisfactory		
ELECTRICAL TDS544A	Tektronix Oscilloscope	16	06-Aug-07	Jane Doe
		13		
		97.98% Satisfactory		
VIBRATION 106	Accelerometer	7	01-Oct-07	Roger Doe
		16		
		99.08% Satisfactory		

Note: The above information is for the period identified only. This report shows what the participant has participated in or is scheduled to participate in. For detailed information regarding specific results, i.e. the En value, Z-scores and IWO Ratings, the Final Report issued to the participant is the official record and should be referenced to for additional evaluation of the participant's results. Preliminary Reports have a specific purpose and should not be used as proof of participations and/or satisfactory ratings. Refer to NAPT's Guidance document or interpreting reports issued by NAPT. Satisfactory/Unsatisfactory rating is based on total number of measurements in the associated block. Questions regarding enrollment or results should be directed to NAPT's Managing Director.